



October 21, 2024

Onondaga CSD
Onondaga CSD
4466 S Onondaga Rd
Nedrow, NY 13120

RE: Project: H S 10/09
Pace Project No.: 70317255

Dear Onondaga CSD:

Enclosed are the analytical results for sample(s) received by the laboratory on October 10, 2024. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Latoya Sobratie
latoya.sobratie@pacelabs.com
516-370-6020
Project Manager

Enclosures



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: H S 10/09

Pace Project No.: 70317255

Pace Analytical Services, LLC - Melville, NY

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: H S 10/09

Pace Project No.: 70317255

Sample: 50 LM #0425.#2		Lab ID: 70317255001		Collected: 10/09/24 05:50		Received: 10/10/24 07:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/21/24 11:58	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: H S 10/09

Pace Project No.: 70317255

QC Batch: 367382

QC Batch Method: EPA 200.8

Analysis Method: EPA 200.8

Analysis Description: 200.8 MET No Prep Drinking Water

Laboratory: Pace Analytical Services - Melville

Associated Lab Samples: 70317255001

METHOD BLANK: 1917712

Matrix: Water

Associated Lab Samples: 70317255001

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/21/24 11:37	

LABORATORY CONTROL SAMPLE: 1917713

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.8	102	85-115	

MATRIX SPIKE SAMPLE: 1917715

Parameter	Units	70317350001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.6	50	40.9	79	70-130	

MATRIX SPIKE SAMPLE: 1917717

Parameter	Units	70317350002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	38.8	76	70-130	

SAMPLE DUPLICATE: 1917714

Parameter	Units	70317350001 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	1.6	1.7	1	

SAMPLE DUPLICATE: 1917716

Parameter	Units	70317350002 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: H S 10/09

Pace Project No.: 70317255

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE


Project: H S 10/09

Pace Project No.: 70317255

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70317255001	50 LM #0425.#2	EPA 200.8	367382		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CGZ/TCO/ : #0M
70317255

Chain of Custody Record

~~Page 7 of 9~~

Shipment Method.

Samples Received Intact: Y N

WO#: 70317255

Client Name:

Ononcsd

Project #

PM: LS2

Due Date: 10/24/24

CLIENT: ONONCSD

Courier: ☐ Fed Ex ☐ UPS ☐ USPS ☐ Client ☐ Commercial ☒ Parcel ☐ Other

Tracking #:

Custody Seal on Cooler/Box Present: ☐ Yes ☒ No Seals intact: ☐ Yes ☒ No Temperature Blank Present: ☐ Yes ☒ No

Packing Material: ☐ Bubble Wrap ☐ Bubble Bags ☐ Ziploc ☒ None ☐ Other Type of Ice: ☒ Wet ☐ Blue ☐ None

Thermometer Used: 1121 Correction Factor: 10.3 ☐ Samples on ice, cooling process has begun

Cooler Temperature (°C): 1.4 Cooler Temperature Corrected (°C): 1.4 Date/Time 5035A kits placed in freezer

Temp should be above freezing to 6.0°C

USDA Regulated Soil (☒ N/A, water sample)

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? ☐ Yes ☒ No

Did samples originate from a foreign source including Hawaii and Puerto Rico? ☐ Yes ☒ No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.

Date and Initials of person examining contents:

ASF 10/11/24

		COMMENTS:
Chain of Custody Present:	☒ Yes ☐ No	1.
Chain of Custody Filled Out:	☒ Yes ☐ No	2.
Chain of Custody Relinquished:	☒ Yes ☐ No	3.
Sampler Name & Signature on COC:	☒ Yes ☐ No ☐ N/A	4.
Samples Arrived within Hold Time:	☒ Yes ☐ No	5.
Short Hold Time Analysis (<72hr):	☐ Yes ☒ No	6.
Rush Turn Around Time Requested	☐ Yes ☒ No	7.
Sufficient Volume: (Triple volume provided for MS/MSD)	☒ Yes ☐ No	8.
Correct Containers Used:	☒ Yes ☐ No	9.
-Pace Containers Used:	☐ Yes ☒ No	
Containers Intact:	☒ Yes ☐ No	10.
Filtered volume received for Dissolved tests	☐ Yes ☐ No ☒ N/A	11. Note: if sediment is visible in the dissolved container.
Sample Labels match COC:	☒ Yes ☐ No	12.
-Includes date/time/ID/Analysis Matrix:	SL ☒ WT ☐ OIL ☐ OTHER	

Date and Initials of person checking preservation:

ASF 10/11/24

All containers needing preservation have been	☒ Yes ☐ No ☐ N/A	13. ☒ HNO ₃ ☐ H ₂ SO ₄ ☐ NaOH ☐ HCl
pH paper Lot # <u>200623</u>		Sample # <u>A11</u>
All containers needing preservation are found to be in compliance with method recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH > 9 Sulfide, ☒ Yes ☐ No ☐ N/A NAOH > 12 Cyanide)		
Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water).		
Per Method, VOA pH is checked after analysis		
Samples checked for dechlorination:	☐ Yes ☐ No ☒ N/A	14. Initial when completed: <u>702</u> Lot # of added preservative: <u>2401-769</u> Date/Time preservative added: <u>10/10/24 17:58</u>
KI starch test strips Lot #		15. Positive for Res. Chlorine? Y N
Residual chlorine strips Lot #		Positive for Sulfide? Y N
SM 4500 CN samples checked for sul	☐ Yes ☐ No ☒ N/A	
Lead Acetate Strips Lot #		
Headspace in ALK Bottle (>6mm):	☐ Yes ☐ No ☒ N/A	16.
Headspace in VOA Vials (>6mm):	☐ Yes ☐ No ☒ N/A	
Trip Blank Present:	☐ Yes ☐ No ☒ N/A	17.
Trip Blank Custody Seals Present	☐ Yes ☐ No ☒ N/A	

Client Notification/ Resolution:

Field Data Required?

Y / N

Person Contacted:

Date/Time:

Comments/ Resolution:

* PM (Project Manager) review (which includes the SCUR) is documented electronically in LIMS.